

Quality Control

## Certificate of Analysis

QC release date: 2021-04-20

**Product:** Creatinine PAP FS

**Product Code:** 11759

Cat. No.: 1 1759 99 10 026

---

**Delivery of kits:** No of kits: 5  
Kit Lot/Los: 60146447  
Kit Expiry: 2022-04-30

---

**Delivery of bulk:** Volume:  
Processing Lot:

---

### Results for:

Reagent 1 Lot No.: 30679  
Expiry: 2022-04-30

Reagent 2 Lot No.: 30679  
Expiry: 2022-04-30

### Examination:

Appearance: Reagent 1: clear, yellow  
Reagent 2: clear, pale brown

Odour: Reagent 1: like detergent  
Reagent 2: no characteristic odour

**pH:** Reagent 1 ( 25 °C ± 1 °C): 8.05  
Reagent 2 ( 25 °C ± 1 °C): 7.98

**Absorbance:** Reagent 1: 0.005  
546 nm / 37 °C R1 + R2: 0.009

**Control sera:** TruLab N TruLab P  
Lot: 27547 Lot: 27548

Test results 1.41 mg/dL 5.24 mg/dL

Target values 1.39 1.08 - 1.70 5.15 4.02 - 6.28

---

**Additional remarks:** Mf-Date: 25.02.2021

This lot meets our specifications.  
This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka