

## CERTIFICADO DE ANÁLISE

### CERTIFICATE OF ANALYSIS

<b>MS ANVISA</b>	<b>80115310042</b>
<b>CÓDIGO / CODE</b>	<b>107095</b>
<b>COD. INTERNO / INTERNAL CODE</b>	<b>1100075K</b>
<b>DESCRIÇÃO / DESCRIPTION</b>	<b>LACTATO R1 3X20ML + R2 1X15ML</b>
	<b>KIT</b>
	LOTE / LOT: 10709521136
	DATA DE FABR./MANUF. DATE: 09/2021
	DATA DE VALIDADE / EXPIRY DATE: 09/2022
	<b>REAGENTE 1 / REAGENT 1</b>
	COD. INTERNO / INTERNAL CODE: 1100020R1K
	LOTE / LOT: 10721321111
	DATA DE VALIDADE / EXPIRY DATE: 09/2022
	<b>REAGENTE 2 / REAGENT 2</b>
	COD. INTERNO / INTERNAL CODE: 1100015R2K
	LOTE / LOT: 10721421111
	DATA DE VALIDADE / EXPIRY DATE: 09/2022

DESCRIÇÃO / DESCRIPTION	ESPECIFICAÇÃO / SPECIFICATION	RESULTADO / RESULTS
APARENCIA/APPEARANCE R1	LIMPIDO, INCOLOR A AMARELADO - CLEAR, COLOURLESS TO YELLOWISH	CONFORME/ACCORDING
APARENCIA/APPEARANCE R2	LIMPIDO, INCOLOR A AMARELADO - CLEAR, COLOURLESS TO YELLOWISH	CONFORME/ACCORDING
ODOR/ODOUR R1	ODOR CARACTERISTICO DETERGENTES - CHARACTERISTIC ODOUR DETERGENT	CONFORME/ACCORDING
ODOR/ODOUR R2	SEM ODOR CARACTERISTICO - NO CHARACTERISTIC ODOUR	CONFORME/ACCORDING
pH (25 oC) R1	8.90 - 9.10	CONFORME/ACCORDING
PH (25 oC) R2	2,50 - 2,70	CONFORME/ACCORDING
pH (25 oC) R1 + R2	8.75 - 9.15	CONFORME/ACCORDING
ABSORBANCIA/ABSORBANCE R1 - 340 nm	0.000 - 0.100	CONFORME/ACCORDING
ABSORBANCIA/ABSORBANCE R2 - 340 nm	0,000 - 0,100	CONFORME/ACCORDING
ABSORBANCIA/ABSORBANCE R1+R2 - 340 nm	0.000 - 0.300	CONFORME/ACCORDING
TRULAB N - LOTE: 28657 (mg/dL)	8.55 - 13.1	9.98
TRULAB P - LOTE: 29696 (mg/dL)	25.1 - 38.5	31.54

OBS.: Este documento foi produzido eletronicamente e é válido sem assinatura / This document has been produced eletronicly and is valid without a signature.

**Monique Brasil Vianna de Sá**  
Controle de Qualidade / Quality Control