

# CERTIFICADO DE ANÁLISE

## CERTIFICATE OF ANALYSIS

<b>MS ANVISA</b>	<b>80115310042</b>
<b>CÓDIGO / CODE</b>	<b>107320</b>
<b>COD. INTERNO / INTERNAL CODE</b>	<b>1100250K</b>
<b>DESCRIÇÃO / DESCRIPTION</b>	<b>LACTATO R1 5X40ML + R2 1X50ML</b>
	<b>KIT</b>
	LOTE / LOT: 1073202095
	DATA DE FABR./MANUF. DATE: 11/2020
	DATA DE VALIDADE / EXPIRY DATE: 02/2022
	<b>REAGENTE 1 / REAGENT 1</b>
	COD. INTERNO / INTERNAL CODE: 1100040R1K
	LOTE / LOT: 1074162094
	DATA DE VALIDADE / EXPIRY DATE: 02/2022
	<b>REAGENTE 2 / REAGENT 2</b>
	COD. INTERNO / INTERNAL CODE: 1100050R2K
	LOTE / LOT: 1074172094
	DATA DE VALIDADE / EXPIRY DATE: 02/2022

<b>DESCRIÇÃO / DESCRIPTION</b>	<b>ESPECIFICAÇÃO / SPECIFICATION</b>	<b>RESULTADO / RESULTS</b>
APARENCIA/APPEARANCE R1	LIMPIDO, INCOLOR A AMARELADO - CLEAR, COLOURLESS TO YELLOWISH	CONFORME/ACCORDING
APARENCIA/APPEARANCE R2	LIMPIDO, INCOLOR A AMARELADO - CLEAR, COLOURLESS TO YELLOWISH	CONFORME/ACCORDING
ODOR/ODOUR R1	ODOR CARACTERISTICO DETERGENTES - CHARACTERISTIC ODOUR DETERGENT	CONFORME/ACCORDING
ODOR/ODOUR R2	SEM ODOR CARACTERISTICO - NO CHARACTERISTIC ODOUR	CONFORME/ACCORDING
pH (25 oC) R1	8.90 - 9.10	CONFORME/ACCORDING
PH (25 oC) R2	2,50 - 2,70	CONFORME/ACCORDING
pH (25 oC) R1 + R2	8.75 - 9.15	CONFORME/ACCORDING
ABSORBANCIA/ABSORBANCE R1 - 340 nm	0.000 - 0.100	CONFORME/ACCORDING
ABSORBANCIA/ABSORBANCE R2 - 340 nm	0,000 - 0,100	CONFORME/ACCORDING
ABSORBANCIA/ABSORBANCE R1+R2 - 340 nm	0.000 - 0.300	CONFORME/ACCORDING
TRULAB N - LOTE: 27547 (mg/dL)	9.01 - 13.80	11.56
TRULAB P - LOTE: 27548 (mg/dL)	24.60 - 37.60	31.42

OBS.: Este documento foi produzido eletronicamente e é válido sem assinatura / This document has been produced eletronically and is valid without a signature.

**Monique Brasil Vianna de Sá**  
Controle de Qualidade / Quality Control