

CERTIFICADO DE ANÁLISE CERTIFICATE OF ANALYSIS

MS ANVISA	80115310042	
CÓDIGO / CODE	107320	
COD. INTERNO / INTERNAL CODE	1100250K	
DESCRIÇÃO / DESCRIPTION	LACTATO R1 5X40ML + R2 1X50ML	
	KIT	
	LOTE / LOT:	10732021100
	DATA DE FABR./MANUF. DATE	07/2021
	DATA DE VALIDADE / EXPIRY DATE:	09/2022
	REAGENTE 1 / REAGENT 1	
	COD. INTERNO / INTERNAL CODE:	1100040R1K
	LOTE / LOT:	1074162199
	DATA DE VALIDADE / EXPIRY DATE:	09/2022
	REAGENTE 2 / REAGENT 2	
	COD. INTERNO / INTERNAL CODE:	1100050R2K
	LOTE / LOT:	1074172199
	DATA DE VALIDADE / EXPIRY DATE:	09/2022

DESCRIÇÃO / DESCRIPTION	ESPECIFICAÇÃO / SPECIFICATION	RESULTADO / RESULTS
APARENCIA/APPEARANCE R1	LIMPIDO, INCOLOR A AMARELADO - CLEAR, COLOURLESS TO YELLOWISH	CONFORME/ACCORDING
APARENCIA/APPEARANCE R2	LIMPIDO, INCOLOR A AMARELADO - CLEAR, COLOURLESS TO YELLOWISH	CONFORME/ACCORDING
ODOR/ODOUR R1	ODOR CARACTERISTICO DETERGENTES - CHARACTERISTIC ODOUR DETERGENT	CONFORME/ACCORDING
ODOR/ODOUR R2	SEM ODOR CARACTERISTICO - NO CHARACTERISTIC ODOUR	CONFORME/ACCORDING
pH (25 oC) R1	8.90 - 9.10	CONFORME/ACCORDING
PH (25 oC) R2	2,50 - 2,70	CONFORME/ACCORDING
pH (25 oC) R1 + R2	8.75 - 9.15	CONFORME/ACCORDING
ABSORBANCIA/ABSORBANCE R1 - 340 nm	0.000 - 0.100	CONFORME/ACCORDING
ABSORBANCIA/ABSORBANCE R2 - 340 nm	0,000 - 0,100	CONFORME/ACCORDING
ABSORBANCIA/ABSORBANCE R1+R2 - 340 nm	0.000 - 0.300	CONFORME/ACCORDING
TRULAB N - LOTE: 28657 (mg/dL)	8.55 - 13.1	10.42
TRULAB P - LOTE: 29696 (mg/dL)	25.1 - 38.5	31.43

OBS.: Este documento foi produzido eletronicamente e é válido sem assinatura / This document has been produced eletronicly and is valid without a signature.

Monique Brasil Vianna de Sá
Controle de Qualidade / Quality Control