

## CERTIFICADO DE ANÁLISE CERTIFICATE OF ANALYSIS

<b>MS ANVISA</b>	<b>80115310042</b>
<b>CÓDIGO / CODE</b>	<b>107320</b>
<b>COD. INTERNO / INTERNAL CODE</b>	<b>1100250K</b>
<b>DESCRIÇÃO / DESCRIPTION</b>	<b>LACTATO R1 5X40ML + R2 1X50ML</b>
	<b>KIT</b>
	LOTE / LOT: 1073202197
	DATA DE FABR./MANUF. DATE: 02/2021
	DATA DE VALIDADE / EXPIRY DATE: 04/2022
	<b>REAGENTE 1 / REAGENT 1</b>
	COD. INTERNO / INTERNAL CODE: 1100040R1K
	LOTE / LOT: 1074162196
	DATA DE VALIDADE / EXPIRY DATE: 04/2022
	<b>REAGENTE 2 / REAGENT 2</b>
	COD. INTERNO / INTERNAL CODE: 1100050R2K
	LOTE / LOT: 1074172196
	DATA DE VALIDADE / EXPIRY DATE: 04/2022

DESCRIÇÃO / DESCRIPTION	ESPECIFICAÇÃO / SPECIFICATION	RESULTADO / RESULTS
APARENCIA/APPEARANCE R1	LIMPIDO, INCOLOR A AMARELADO - CLEAR, COLOURLESS TO YELLOWISH	Conforme/According
APARENCIA/APPEARANCE R2	LIMPIDO, INCOLOR A AMARELADO - CLEAR, COLOURLESS TO YELLOWISH	Conforme/According
ODOR/ODOUR R1	ODOR CARACTERISTICO DETERGENTES - CHARACTERISTIC ODOUR DETERGENT	Conforme/According
ODOR/ODOUR R2	SEM ODOR CARACTERISTICO - NO CHARACTERISTIC ODOUR	Conforme/According
pH (25 oC) R1	8.90 - 9.10	Conforme/According
PH (25 oC) R2	2,50 - 2,70	Conforme/According
pH (25 oC) R1 + R2	8.75 - 9.15	Conforme/According
ABSORBANCIA/ABSORBANCE R1 - 340 nm	0.000 - 0.100	Conforme/According
ABSORBANCIA/ABSORBANCE R2 - 340 nm	0,000 - 0,100	Conforme/According
ABSORBANCIA/ABSORBANCE R1+R2 - 340 nm	0.000 - 0.300	Conforme/According
TRULAB N - LOTE: 27547 (mg/dL)	9.01 - 13.80	11.87
TRULAB P - LOTE: 27548 (mg/dL)	24.60 - 37.60	31.54

OBS.: Este documento foi produzido eletronicamente e é válido sem assinatura / This document has been produced electronically and is valid without a signature.

**Monique Brasil Vianna de Sá**  
Controle de Qualidade / Quality Control