

Quality Control

## Certificate of Analysis

QC release date: 2020-04-16

**Product:** Total Protein UC FS  
**Product Code:** 10210

Cat. No.: 1 0210 99 10 021

---

**Delivery of kits:** No of kits: 35  
Kit Lot/Los: 60140346  
Kit Expiry: 2022-03

---

**Delivery of bulk:** Volume:  
Processing Lot:

---

### Results for:

Reagent Lot No.: 29180  
Expiry: 2022-03-24

### Examination:

Visual check: Reagent : clear, orange-red  
Smell Reagent : no characteristic odour

**pH:** Reagent ( 25 °C ± 1 °C): 2.54

**Absorbance:** Reagent : 0.105  
600 nm/ 15-25°C

<b>Control sera:</b>	TruLab Urin 1	TruLab Urin 2
	Lot: 26478	Lot: 26479
Fresh (at 37°C)	0.084 g/L	0.640 g/L
Target values	0.083 0.063 - 0.103	0.624 0.474 - 0.774

---

**Additional remarks:** Mf-Date: 24.03.2020

This lot meets our specifications.  
This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka