

Quality Control

Certificate of Analysis

QC release date: 2021-05-19

Product: Total Protein UC FS
Product Code: 10210

Cat. No.: 1 0210 99 10 021

Delivery of kits: No of kits: 24
Kit Lot/Los: 60145722
Kit Expiry: 2023-01-29

Delivery of bulk: Volume:
Processing Lot:

Results for:

Reagent Lot No.: 30809
Expiry: 2023-04-30

Examination:

Appearance: Reagent : clear, orange-red
Odour: Reagent : no characteristic odour
pH: Reagent (25 °C ± 1 °C): 2.46
Absorbance: Reagent : 0.105
600 nm/ 15-25°C

Control sera:	TruLab Urine 1	TruLab Urine 2
	Lot: 30530	Lot: 30531
Fresh (at 37°C)	0.078 g/L	0.665 g/L
Target values	0.075 0.057 - 0.093	0.607 0.461 - 0.753

Additional remarks: Mf-Date: 30.04.2021

This lot meets our specifications.
This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka